



The Children's Hospital & The Institute of Child Health
Ferozpur Road, Lahore Phone # (92) (42) 99230901-23 Fax # 99231560



No. P.C./ 13671 /CH&ICH

Dated: 13-03- 2018


To

✓
Mr. Abdul Wahab
I/C Programmer (I.T.)
Health Department, Civil Secretariat,
Lahore.

Subject: - **REQUEST FOR UPLOADING THE TECHNICAL EVALUATION FOR THE PROCUREMENT OF ELECTRO MEDICAL EQUIPMENT FOR ESTABLISHMENT OF PAEDIATRIC CARDIOLOGY AND CARDIC SURGERY AND ALLIED SPECIALITIES IN THE CHILDREN'S HOSPITAL, LAHORE FOR THE YEAR 2017-2018 ON PPRA WEBSITE/HEALTH DEPARTMENT WEBSITE**

Please find enclosed herewith bids technical evaluation reports for the procurement of **Electro Medical Equipment for Establishment of Paediatric Cardiology & Cardic Surgery and Allied Specialities** for the year **2017-2018** in The Children's Hospital, Lahore and needs to be uploaded in the Health Department / PPRA website.


Your positive response will be highly appreciated.


Prof. Dr. Muhammad Saleem
(FCPS, MME-HPE)
Professor of Paediatric Surgery
Medical Director

No. P.C./ /CH&ICH

Dated _____ 2018

A copy is forwarded for information to Additional Secretary (Development), Government of the Punjab, Specialized Healthcare & Medical Education Department, Lahore.


Prof. Dr. Muhammad Saleem
(FCPS, MME-HPE)
Professor of Paediatric Surgery
Medical Director



The Children's Hospital & The Institute of Child Health

Ferozpur Road, Lahore

Phone No. +92 42 99230901 - 23

Technical Evaluation

Package No. / Bid No. : 5369 / 02 / CH & ICH

Name of the Equipment:

a) E-Journals Subscription
b) Journals (Peads Cardiology,
Cardiology in the Young)

Required Quantity:

a) = 5 & b) 6

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by Purchase Committee

Sr. #	Evaluation Parameters	Allied Book Store
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security / CDR (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

Package No. / Bid No. : 5369 / 02 / CH & ICH

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by Technical Committee

Sr. #	Evaluation Parameters	Allied Book Store
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	N/A
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	N/A
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	N/A
04	Bidder Past Performance	N/A
05	Availability of relevant Tools and Testing / Calibration Equipment	N/A
06	Warranty	N/A
Remarks:		Responsive

Package No. / Bid No. : 5369 / 02 / CH & ICH

PART - III		
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)		
Evaluated by Technical Committee		
Req. Item #	Required Item Name	Allied Book Store
2	Docking Station for 200 Syringe Pumps	
	Brand	N/A
Sr. #	Quoted Model	N/A
01	Country of Manufacturer	N/A
02	Country of Origin of Product	N/A
03	Compliance with defined Quality Standards	N/A
04	Specifications Compliance Feature Wise	N/A
05	Technical Eligibility of Product	N/A
06	Technical Eligibility of Firm	Accepted
Technical Status		Responsive

Ahmed Maqsood
Bio Medical Engineer

Muhammad Umair Kabeer
Bio Medical Engineer

Dr. Khalid Hameed Zaighum
DMS (Purchase)

Dr. Osaid Qureshi
Assistant Prof. of Paed Cardiology

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Paed
Cardiovascular Surgery

Dr. Syed Najam Haider
Associated Prof. of Paed Cardiology



The Children's Hospital & The Institute of Child Health

Ferozpur Road, Lahore

Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : 5369 / 05 / CH & ICH

Name of the Equipment:

Docking Station for 200
Syringe Pumps

Required Quantity:

Set

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by Purchase Committee

Sr. #	Evaluation Parameters	M/s Medi Urge
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security / CDR (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

Package No. / Bid No. : 5369 / 05 / CH & ICH

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by Technical Committee

Sr. #	Evaluation Parameters	M/s Medi Urge
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance	Satisfactory
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive

Package No. / Bid No. : 5369 / 05 / CH & ICH

PART - III		
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)		
<i>Evaluated by Technical Committee</i>		
Req. Item #	Required Item Name	M/s Medi Urge
5	Docking Station for 200 Syringe Pumps	Fresenius
	Brand	
Sr. #	Quoted Model	Link 4 Agilia Int.
01	Country of Manufacturer	France
02	Country of Origin of Product	France
03	Compliance with defined Quality Standards	Compliance
04	Specifications Compliance Feature Wise	Accepted
05	Technical Eligibility of Product	Accepted
06	Technical Eligibility of Firm	Accepted
Technical Status		Responsive

Ahmed Maqsood
Bio Medical Engineer

Muhammad Umair Kabeer
Bio Medical Engineer

Dr. Khalid Hameed Zaighum
DMS (Purchase)

Dr. Osaid Qureshi
Assistant Prof. of Paed Cardiology

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Paed
Cardiovascular Surgery

Dr. Syed Najam Haider
Associated Prof. of Paed Cardiology



The Children's Hospital & The Institute of Child Health
Ferozpur Road, Lahore
Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : 5369 / 06 / CH & ICH

Name of the Equipment:

Head Light with Light Source

Required Quantity:

1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by Purchase Committee

Sr. #	Evaluation Parameters	Medi Urge	Dymedix	Allmed Solution
01	Complete Package / Tender (Mandatory)	Yes	No	No
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	No	Yes
04	Bid Security / CDR (Mandatory)	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes
Remarks:		Responsive	Non - Responsive	Non - Responsive

Package No. / Bid No. : 5369 / 06 / CH & ICH

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by Technical Committee

Sr. #	Evaluation Parameters	Medi Urge	Dymedix	Allmed Solution
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	N/A	No	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	N/A	No	Yes
04	Bidder Past Performance	Satisfactory	Not Found	Not Found
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes
Remarks:		Responsive	Non - Responsive	Non - Responsive

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by Technical Committee

Req. Item #	Required Item Name	Medi Urge	Dymedix	Allmed Solution
6	Head Light with Light Source	Luxtec	Ecleris	Karl Storz
	Brand			
Sr. #	Quoted Model	MLX 300 Watt	Head Lux	Head Light KS 70
01	Country of Manufacturer	USA	USA	Germany
02	Country of Origin of Product	USA	USA	Germany
03	Compliance with defined Quality Standards	Compliance	Compliance	Compliance
04	Specifications Compliance Feature Wise	Accepted	1) Affidivit not attached 2) Equipment Cart not quoted 3) No past performance in Teaching Hospitals of Punjab 4) Poor performance history in CH&ICH, Lhr	1) No light source quoted (Incomplete Package) 2) No past performance in Teaching Hospitals of Punjab
05	Technical Eligibility of Product	Accepted	Not Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Not Accepted	Accepted
Technical Status		Responsive	Non - Responsive	Non - Responsive

Ahmed Maqsood
Bio Medical Engineer

Muhammad Umair Kabeer
Bio Medical Engineer

Dr. Khalid Hameed Zaighum
DMS (Purchase)

Dr. Osaid Qureshi
Assistant Prof. of Paed Cardiology

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Paed Cardiovascular Surgery

Dr. Syed Najam Haider
Associated Prof. of Paed Cardiology



The Children's Hospital & The Institute of Child Health
Ferozpur Road, Lahore
Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : 5369 / 09 / CH & ICH

Name of the Equipment:

Cardiac ICU Bed with Overhead Table (Imported)

Required Quantity:

6

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by Purchase Committee

Sr. #	Evaluation Parameters	KASBN	Medequips	Medi Urge
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes
04	Bid Security / CDR (Mandatory)	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive

Package No. / Bid No. : 5369 / 09 / CH & ICH

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by Technical Committee

Sr. #	Evaluation Parameters	KASBN	Medequips	Medi Urge
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes
04	Bidder Past Performance	Satisfactory	Satisfactory	Satisfactory
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive

Package No. / Bid No. : 5369 / 09 / CH & ICH

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by Technical Committee

Req. Item #	Required Item Name	KASBN	Medequips	Medi Urge
9	Cardiac ICU Bed with Overhead Table (Imported)			
	Brand	Industrias Hidraulicas PADRO	Paramount Bed	Linetspols.r.o
Sr. #	Quoted Model	Newcare V2	Qualitas Plus Series	ELEGNZA2
01	Country of Manufacturer	Spain	Japan	Czech Republic
02	Country of Origin of Product	Spain	Indonesia	Czech Republic
03	Compliance with defined Quality Standards	Compliance	Compliance	Compliance
04	Specifications Compliance Feature Wise	Subject to Demo	Subject to Demo	Subject to Demo
05	Technical Eligibility of Product	Accepted	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted	Accepted
Technical Status		Responsive	Responsive	Responsive

Ahmed Maqsood
Bio Medical Engineer

Muhammad Umair Kabeer
Bio Medical Engineer

Dr. Khalid Hameed Zaighum
DMS (Purchase)

Dr. Osaid Qureshi
Assistant Prof. of Paed Cardiology

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Paed
Cardiovascular Surgery

Dr. Syed Najam Haider
Associated Prof. of Paed Cardiology



The Children's Hospital & The Institute of Child Health
Ferozpur Road, Lahore
Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : 5369 / 20 / CH & ICH

Name of the Equipment:

Cochlear Implant

Required Quantity:

10

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by Purchase Committee

Sr. #	Evaluation Parameters	MED-EL Pakistan	Hearing Clinic	Heartec
01	Complete Package / Tender (Mandatory)	Yes	No	No
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes
04	Bid Security / CDR (Mandatory)	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes
Remarks:		Responsive	Non - Responsive	Non - Responsive

Package No. / Bid No. : 5369 / 20 / CH & ICH

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by Technical Committee

Sr. #	Evaluation Parameters	MED-EL Pakistan	Hearing Clinic	Heartec
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes
04	Bidder Past Performance	Satisfactory	No	No
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes
Remarks:		Responsive	Non - Responsive	Non - Responsive

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by Technical Committee

Req. Item #	Required Item Name	MED-EL Pakistan	Hearing Clinic	Heartec
20	Cochlear Implant			
	Brand	MED-EL	Advance Bionics	Oticon Medical / Neurelec
Sr. #	Quoted Model	Mi1210 & Opus 2	Not Mentioned	Digisonic SP with SAPHYR Neo Speech Processor
01	Country of Manufacturer	Austria	USA	France
02	Country of Origin of Product	Austria	USA & Switzerland	France
03	Compliance with defined Quality Standards	Compliance	Compliance	Compliance
04	Specifications Compliance Feature Wise	Accepted	1) Incomplete Bid (No signed & stamped bidding docs attached) 2) Offered model not mentioned in technical offer 3) No past experience in any teaching hospital of Punjab	1) No past experience in any teaching hospital of Punjab 2) Low Stimulation rate, i.e 24K PPS instead of 40K 3) Pulse Amplitude not mentioned & Low Pulse width, i.e. upto 250µs instead 400µs
05	Technical Eligibility of Product	Accepted	Not Accepted	Not Accepted
06	Technical Eligibility of Firm	Accepted	Not Accepted	Not Accepted
Technical Status		Responsive	Non - Responsive	Non - Responsive

Ahmed Maqsood
Bio Medical Engineer

Muhammad Umair Kabeer
Bio Medical Engineer

Dr. Khalid Hameed Zaighum
DMS (Purchase)

Dr. Samra
Senior Medical Officer Audiology
Department

Dr. Imran Saeed
Associated Prof. of Paed ENT